


## Instruction Manual

---

# HI 9813-5 & HI 9813-6 Portable pH/EC/TDS/°C Meters



Dear Customer,  
Thank you for choosing a Hanna product.  
Please read this instruction manual carefully before using the meter.  
This manual will provide you with the necessary information for correct use of the instrument, as well as a precise idea of its versatility.  
If you need additional technical information, do not hesitate to e-mail us at [tech@hannainst.com](mailto:tech@hannainst.com).

## WARRANTY

HI 9813-5 and HI 9813-6 are guaranteed for two years against defects in workmanship and materials when used for their intended purpose and maintained according to instructions. Electrodes and probes are guaranteed for six months. This warranty is limited to repair or replacement free of charge.

Damage due to accidents, misuse, tampering or lack of prescribed maintenance is not covered.

If service is required, contact the dealer from whom you purchased the instrument. If under warranty, report the model number, date of purchase, serial number and the nature of the problem. If the repair is not covered by the warranty, you will be notified of the charges incurred. If the instrument is to be returned to Hanna Instruments, first obtain a Returned Goods Authorization number from the Technical Service department and then send it with shipping costs prepaid. When shipping any instrument, make sure it is properly packed for complete protection.

## TABLE OF CONTENTS

WARRANTY .....	2
PRELIMINARY EXAMINATION .....	3
GENERAL DESCRIPTION .....	3
FUNCTIONAL DESCRIPTION OF HI 9813-5 .....	4
FUNCTIONAL DESCRIPTION OF HI 9813-6 .....	5
SPECIFICATIONS .....	6
OPERATIONAL GUIDE .....	7
pH CALIBRATION .....	9
pH VALUES AT VARIOUS TEMPERATURES .....	10
EC/TDS CALIBRATION .....	11
BATTERY REPLACEMENT .....	12
PROBE MAINTENANCE .....	12
ACCESSORIES .....	14

## PRELIMINARY EXAMINATION

Remove the instrument from the packing material and examine it carefully to make sure that no damage has occurred during shipment. If noticeable damage is found, notify your dealer.

Each meter is supplied with:

- Combination probe with 8-pin DIN connector & 1 m (3.3') cable:
  - HI 1285-5, pH/EC/TDS/°C probe, for HI 9813-5
  - HI 1285-6, pH/EC/TDS/°C probe with Cal-Check™ for HI 9813-6
- HI 70007 pH 7.01 sachet, 1 pc
- HI 70031 1413  $\mu\text{S}/\text{cm}$  (1.413 mS/cm) sachet, 1 pc
- HI 70442 1500 ppm sachet, 1 pc
- HI 700661 cleaning solution sachet, 2 pcs
- HI 50021 check solution sachet, 2 pcs (for HI 9813-6 only)
- Instruction manual
- 1 x 9V alkaline battery.

**Note:** Save all packing material until you are sure that the instrument functions correctly. Any defective items must be returned in the original packing together with the supplied accessories.

## GENERAL DESCRIPTION


HI 9813-5 and HI 9813-6 are complete, versatile and splash-proof portable combination meters designed with the utmost precision and simplicity. Both models provide measurements of pH, EC and TDS ranges, which can be easily selected using a membrane keypad on the front panel.

Conductivity measurements are automatically compensated for temperature changes with a built-in temperature sensor. The temperature coefficient is fixed at 2%/°C.

- HI 9813-5 is a pH/EC/TDS meter designed for simplicity of use in taking pH, mS/cm, ppm and temperature in the °C scale measurements. Suited for hydroponics, greenhouses, farming and ground water applications.
- HI 9813-6 is the most advanced model, including the Cal-Check™ feature, which allows the user to easily check the probe calibration status at any time.


*All rights are reserved. Reproduction in whole or in part is prohibited without the written consent of the copyright owner.*

## FUNCTIONAL DESCRIPTION OF HI 9813-5



- 1) 8-pin DIN connector for probe
- 2) Liquid Crystal Display
- 3) pH range selection key
- 4) mS/cm (EC) range selection key
- 5) ppm (TDS) selection key
- 6) pH offset calibration knob
- 7) EC/TDS calibration knob
- 8) °C (Temperature) selection key
- 9) On/Off key

## FUNCTIONAL DESCRIPTION OF HI 9813-6



- 1) 8-pin DIN connector for probe
- 2) Liquid Crystal Display
- 3) pH range selection key
- 4) mS/cm (EC) range selection key
- 5) ppm (TDS) range selection key
- 6) pH offset calibration knob
- 7) EC/TDS calibration knob
- 8) °C (Temperature) selection key
- 9) Check key
- 10) On/Off key

## SPECIFICATIONS

<b>Range</b>	0.0 to 14.0 pH 0.00 to 4.00 mS/cm 0 to 1999 ppm 0.0 to 60.0 °C
<b>Resolution</b>	0.1 pH 0.01 mS/cm 1 ppm 0.1 °C
<b>Accuracy (@ 20 °C/68 °F)</b>	±0.1 pH ±2% f.s. mS/cm ±2% f.s. ppm ±0.5 °C
<b>Typical EMC Deviation</b>	±0.1 pH ±2% f.s. mS/cm ±2% f.s. ppm ±0.5 °C
<b>Conversion Factor</b>	variable, 0.56 to 0.78
<b>pH Calibration</b>	Manual, 1-point through offset trimmer
<b>EC/TDS Calibration</b>	Manual, 1-point through slope trimmer
<b>EC/TDS Temperature Compensation</b>	Automatic from 0 to 50 °C (32 to 122 °F) with $\beta=2\%/^{\circ}\text{C}$
<b>Probe (included)</b>	HI 1285-5 for HI 9813-5 HI 1285-6 for HI 9813-6
<b>Battery Type</b>	1 x 9V alkaline
<b>Battery Life</b>	Approx. 150 hours of continuous use
<b>Environment</b>	0 to 50 °C (32 to 122 °F); 100% RH
<b>Dimensions</b>	144.6 x 79.5 x 37 mm (5.7 x 3.1 x 1.5 ")
<b>Weight</b>	230 g (8.1 oz.)

# OPERATIONAL GUIDE

## INITIAL PREPARATION

Each meter is supplied complete with a 9V battery. Remove the battery compartment cover on the back of the meter and install the battery while observing its polarity.

Connect the probe to the DIN socket on the top of the meter by aligning the pins with the socket and pushing in the plug.

Always remove the probe protective cap before taking any measurements, and make sure the meter has been calibrated.



Turn the meter on by pressing the On/Off key.


## TAKING pH MEASUREMENTS

If the probe has been left dry, soak the tip in HI 70300 storage solution for 30 minutes to reactivate it.

- To take a pH measurement simply submerge the tip (4cm/1½") of the probe into the sample to be tested.
- Select the pH mode.



- Stir briefly and wait a couple of minutes for the reading to adjust and stabilize. The display shows the pH value.



- If measurements are taken in different samples successively, it is recommended to rinse (clean) the probe thoroughly to eliminate cross-contamination. After cleaning, it is recommended to rinse the probe with some of the next sample to be measured.

## TAKING EC/TDS MEASUREMENTS



- Immerse the tip of the probe (4cm/1½") into the sample to be tested. If possible, use plastic beakers or containers to minimize any EMC interference.
- Tap the probe lightly on the bottom of the beaker to remove any air bubbles which may be trapped inside the tip.



- Select the appropriate measurement range (EC or TDS).



- Wait for a couple of minutes for the temperature sensor to reach thermal equilibrium. The display then shows the measurement automatically compensated for temperature with the appropriate following indication: “mS” tag indicates the meter is in EC mode, while “ppm” tag indicates the meter is in TDS mode.



### TAKING TEMPERATURE MEASUREMENTS

- Submerge the tip (4cm/1½") of the probe into the sample to be tested.

- Select the °C mode.




- Stir briefly and wait a couple of minutes for the reading to adjust and stabilize. The display shows the temperature value.



- Notes:**
- If the display shows only a “1” on the far left hand side, the reading is out of range.

- It is recommended to clean often the probe with **HI 700661** Cleaning Solution.



- After measurements have been completed, the instrument should be switched off, and the probe cleaned and covered with the protective cap.



## pH CALIBRATION

For greatest accuracy, frequent calibration of the instrument is recommended. The instrument should be recalibrated for pH:

- Whenever the electrode is replaced.
- At least once a month.
- After testing aggressive chemicals.
- Where extreme accuracy is required.
- Whenever the CAL-CHECK test fails (HI 9813-6 only).

### PREPARATION

Pour small quantities of pH 7.01 (HI 7007) or pH 4.01 (HI 7004) or pH 10.01 (HI 7010) solution into a clean beaker.

To obtain accurate readings, use pH 7.01 (HI 7007) if you are going to measure neutral or close to neutral samples, pH 4.01 (HI 7004) if you are going to measure acidic samples or pH 10.01 (HI 7010) for alkaline measurements.

If you need to calibrate to NBS standards, use pH 6.86 (HI 7006) instead of pH 7.01 and pH 9.18 (HI 7009) instead of pH 10.01.

### PROCEDURE

- Connect the probe and switch the meter on, then press the **pH** key to display pH measurement.
- Remove the protective cap from the probe, rinse and immerse it in the buffer and stir gently. Wait a couple of minutes for the reading to stabilize.
- Take the temperature of the buffer solution, e.g. 10.0°C, as follow: select the °C mode and read the displayed value.
- Adjust the pH calibration knob until the LCD shows the pH value at the above temperature (see the pH versus temperature chart).
- The pH calibration is now complete.



- Notes:**
- The probe should be submerged approximately 4 cm (1½") into the solution. The thermometer has to be located close to the probe.
  - If turning the knob the needed value can not be reached, clean the probe (see the "Probe Maintenance" section). If also after the probe cleaning the value can not be reached, replace the probe.

## pH VALUES AT VARIOUS TEMPERATURES

For temperature compensation during calibration, please refer to the following chart.

TEMP		pH VALUES				
°C	°F	4.01	6.86	7.01	9.18	10.01
0	32	4.01	6.98	7.13	9.46	10.32
5	41	4.00	6.95	7.10	9.39	10.24
10	50	4.00	6.92	7.07	9.33	10.18
15	59	4.00	6.90	7.05	9.27	10.12
20	68	4.00	6.88	7.03	9.22	10.06
25	77	4.01	6.86	7.01	9.18	10.01
30	86	4.02	6.85	7.00	9.14	9.96
35	95	4.03	6.84	6.99	9.11	9.92
40	104	4.04	6.84	6.98	9.07	9.88
45	113	4.05	6.83	6.98	9.04	9.85
50	122	4.06	6.83	6.98	9.01	9.82
55	131	4.08	6.84	6.98	8.99	9.79
60	140	4.09	6.84	6.98	8.97	9.77
65	149	4.11	6.84	6.99	8.95	9.76
70	158	4.12	6.85	6.99	8.93	9.75

For instance, if the buffer temperature is 25°C, the display should show pH 4.0 or 7.0 or 10.0.

If the buffer temperature is 10°C, the display should show pH 4.0 or 7.0 or 10.1.

## EC/TDS CALIBRATION

### Accessories needed:

- Use **HI 70031** ( $1413 \mu\text{S}/\text{cm} = 1.413 \text{ mS}/\text{cm}$ ) EC calibration solution or **HI 70442** (1500 ppm) TDS calibration solution.

**Note:** The conversion between EC and TDS is made by a built-in circuit, hence it is requested to calibrate the meter only in EC or TDS range. The other range is thus automatically calibrated.

### PROCEDURE

- Pour approximately 4 cm (1½") of a conductivity calibration solution (e.g. **HI 7031**) into a beaker. If possible, use plastic beaker to minimize any EMC interference.
- Immerse the probe in the solution.
- Wait for a couple of minutes for thermal equilibrium to be reached.
- Tap the probe on the bottom, then shake it lightly while rotating to make sure no air bubbles remain trapped inside the probe.
- Press the **mS/cm** (or **ppm**) key.
- Turn the EC/TDS calibration knob until the display shows the EC or TDS reading at 25°C.



## BATTERY REPLACEMENT

These meters are powered by a 9V battery that is located on the rear of the instrument.

When battery symbol is empty a low battery condition is indicated.




When the low battery indication appears, only a few hours of working time remains.

It is recommended to replace the battery immediately.

When the battery level is so low that it may cause unreliable measurements, the meter turns off.

Battery replacement must only take place in a nonhazardous area using a 9V alkaline battery.

Unscrew the three screws on the rear of the meter, remove the battery compartment cover and replace the 9V battery with a new one while observing its polarity.



Make sure the battery contacts are tight and secure before replacing the cover.

## PROBE MAINTENANCE

### PERIODIC MAINTENANCE

Inspect the probe and the cable. The cable used for the connection to the meter must be intact and there must be no points of broken insulation on the cable or cracks on the probe stem or bulb.

Connector must be perfectly clean and dry. If any scratches or cracks are present, replace the electrode. Rinse off any salt deposits with water.

## CLEANING PROCEDURE

For better accuracy in measurements and to ensure a good performance of the probe, a frequent cleaning is recommended.

For this purpose, soak it in Hanna HI 700661 Cleaning Solution for 5 minutes.

**Note:** For particular dirty (as for example protein, oil or grease) see the "Accessories" section for Hanna specific solutions.


HI 9813-6 only:


### CAL-CHECK™ FEATURE / CLEAN & CAL PROCEDURE

The unique Cal-Check™ feature allows the user to check the meter calibration status at any time. Simply follow the below procedure for a quick and easy operation.

- Rinse the probe with water.
- Immerse the probe in the HI 50021 check solution.


**Note:** If the reading is around pH 4.0, the probe is broken and has to be replaced.

- Press the **Check** key. 
- If the meter is calibrated "Probe is OK" message is displayed.



- If cleaning is needed "Clean Probe and Calibrate" message is displayed:

1. immerse the probe in the HI 700661 cleaning solution (or in one of the available specific solutions - see "Accessories" section) for 5 minutes, then rinse it with tap water.
2. immerse again in the HI 50021 check solution and press the CHECK key.
3. if the message "Probe is OK" is displayed, the meter does not need calibration.
4. if the message "Clean Probe and Calibrate" is displayed again, calibration is needed (see "pH Calibration" section).



**Notes:** • If it is not possible to calibrate, the probe has to be replaced with a new one.

- For field applications, it is always recommended to keep a spare probe handy. When anomalies are not resolved with simple maintenance, change the probe and recalibrate the meter.

## ACCESSORIES

### PROBES

- HI 1285-5 Combination, amplified pH/EC/TDS/temperature probe with built-in temperature sensor, 8-pin DIN connector and 1 m (3.3') cable
- HI 1285-6 Combination, amplified pH/EC/TDS/temperature probe with Cal-Check feature, built-in temperature sensor, 8-pin DIN connector and 1 m (3.3') cable

### pH BUFFER SOLUTIONS

- HI 7004L pH 4.01 buffer solution, 500 mL bottle
- HI 7006L pH 6.86 buffer solution, 500 mL bottle
- HI 7007L pH 7.01 buffer solution, 500 mL bottle
- HI 7009L pH 9.18 buffer solution, 500 mL bottle
- HI 7010L pH 10.01 buffer solution, 500 mL bottle

### CONDUCTIVITY & TDS CALIBRATION SOLUTIONS

- HI 7031L 1413  $\mu$ S/cm solution, 500 mL bottle
- HI 7442L 1500 ppm (mg/L) solution, 500 mL bottle

### OTHER SOLUTIONS

- HI 700661P Cleaning solution, 20 mL sachet (25 pcs)
- HI 50021P Check solution, 20 mL sachet (25 pcs)
- HI 70300L Storage solution, 500 mL bottle
- HI 7073L Protein cleaning solution, 500 mL bottle
- HI 7074L Inorganic cleaning solution, 500 mL bottle
- HI 7077L Oil & Fat cleaning solution, 500 mL bottle

### OTHER ACCESSORIES

- HI 710007 Shockproof rubber boot, blue
- HI 710008 Shockproof rubber boot, orange
- HI 710022 Blue protective case

### **RECOMMENDATIONS FOR USERS**

Before using these products, make sure they are entirely suitable for the environment in which they are used.

Operation of these instruments in residential areas could cause unacceptable interferences to radio and TV equipment, requiring the operator to follow all necessary steps to correct interferences.

The glass bulb at the end of the electrode is sensitive to electrostatic discharges. Avoid touching this glass bulb at all times.

During operation, ESD wrist straps should be worn to avoid possible damage to the electrode by electrostatic discharges.

Any variation introduced by the user to the supplied equipment may degrade the instrument's EMC performance.

To avoid electrical shock, do not use these instruments when voltages at the measurement surface exceed 24 VAC or 60 VDC.

To avoid damages or burns, do not perform any measurement in microwave ovens.

Hanna Instruments reserves the right to modify the design, construction and appearance of its products without advance notice.



**Hanna Instruments Inc.**  
Highland Industrial Park  
584 Park East Drive  
Woonsocket, RI 02895 USA

**Technical Support for Customers**  
Tel. (800) 426 6287  
Fax (401) 765 7575  
E-mail [tech@hannainst.com](mailto:tech@hannainst.com)  
[www.hannainst.com](http://www.hannainst.com)

**Local Sales and Customer Service Office**

